

FACT SHEET FOR HEALTHCARE PERSONNEL

ASP STERRAD Sterilization System for Decontaminating Compatible N95 Respirators April 11, 2020

Coronavirus
Disease 2019
(COVID-19)

You have been given a **N95 or N95-equivalent respirator** (“compatible N95 respirator”) that has been decontaminated using a sterilization system **for single-user reuse by healthcare personnel in a healthcare setting** to help prevent exposure to pathogenic biologic airborne particulates during the COVID-19 pandemic.

This Fact Sheet informs you of the significant known and potential risks and benefits of the emergency use of decontaminated, compatible N95 respirators. These compatible N95 respirators have been decontaminated using one of three Advanced Sterilization Products STERRAD Sterilization Systems, using the STERRAD Sterilization Cycles for N95 respirator decontamination: STERRAD 100S Cycle, STERRAD NX Standard Cycle, or STERRAD 100NX Express Cycle (hereafter referred to as “**decontaminated N95 respirators**” and “**ASP STERRAD Sterilization System**” throughout this Fact Sheet).

Decontaminated N95 respirators that have been decontaminated using an ASP STERRAD Sterilization System are authorized for single-user reuse by healthcare personnel in a healthcare setting during the COVID-19 pandemic.

Whether or not you use a respirator, always follow infection control measures: wash hands, cover coughs and sneezes, stay home if you may be sick.

What are the symptoms of COVID-19?

Many patients with confirmed COVID-19 have developed fever and/or symptoms of acute respiratory illness (e.g., cough, difficulty breathing). However, limited information is currently available to characterize the full spectrum of clinical illness associated with COVID-19. Based on what is known about the virus that causes COVID-19, signs and symptoms may appear any time from 2 to 14 days after exposure to the virus. Based on preliminary data, the median incubation period is approximately 5 days, but may range 2-14 days.

Public health officials have identified cases of COVID-19 infection throughout the world, including the United States, which may pose risks for public health. Please check the CDC webpage for the most up to date information.

What do I need to know about the emergency use of decontaminated N95 respirators?

- The ASP STERRAD Sterilization System has been authorized for emergency use to decontaminate compatible N95 respirators for single-user reuse by healthcare personnel during the COVID-19 pandemic to prevent exposure to pathogenic airborne particulates.
 - Compatible N95 or N95-equivalent respirators are those that are compatible with vaporized hydrogen peroxide (VH₂O₂) gas sterilization and can safely and effectively be processed two times after initial use without detrimentally impacting mask form, fit, or function.
 - Cellulose-based materials are incompatible with the ASP STERRAD’s hydrogen peroxide-based sterilization system.
- **Preparing compatible N95 respirators for decontamination:**
 - ✓ Place compatible N95 respirators after use into a compatible sterilization pouch identified for use in vaporized hydrogen peroxide, such as a Tyvek[®] pouch with STERRAD Chemical Indicator
 - ✓ Write name and/or other identifier using a permanent marker so the respirator may be returned after successful decontamination
 - ✓ Place a tick mark on respirator each time a respirator is prepared for decontamination
 - ✓ Post-decontamination, open the pouches and allow the decontaminated N95 respirators to aerate for one hour prior to usage
 - ✓ **Discard if decontaminated more than 2 times** or if visibly soiled or damaged
- **Use of decontaminated N95 respirators:**
 - ✓ Decontaminated N95 respirators are not sterile
 - ✓ Inspect respirators after each use prior to submission for decontamination
 - ✓ If decontaminated N95 respirators are soiled or damaged, they should be discarded

Report Adverse events, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 (<https://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home>) or by calling **1-800-FDA-1088**

FACT SHEET FOR HEALTHCARE PERSONNEL

ASP STERRAD Sterilization System for Decontaminating Compatible N95 Respirators April 11, 2020

Coronavirus
Disease 2019
(COVID-19)

- ✓ Cellulose-based materials are incompatible with the STERRAD Sterilization System
- ✓ Report problems with decontaminated N95 respirators to your healthcare facility

It is strongly recommended to maintain chain of custody on the respirator to minimize the risk of cross-contamination

- **Monitor healthcare personnel for signs and symptoms** of potential infection with SARS-CoV-2 or other respiratory infection for up to and including 14 days after last contact with the SARS-CoV-2 virus and related material, and promptly report such information to Advanced Sterilization Products.
- **Report damage or discoloration** observed upon receipt of the decontaminated, compatible N95 respirators, and potential exposure of healthcare personnel from breaks in or other damage to or degradation of the decontaminated compatible N95 respirators.

Use appropriate personal protective equipment (PPE) when caring for individuals suspected of having COVID-19 as outlined in the CDC webpages, including *Interim Infection Prevention and Control Recommendations for Patients with Confirmed Coronavirus Disease 2019 (COVID-19) or Persons Under Investigation for COVID-19 in Healthcare Settings, Infection Control, and FAQ about PPE*.

Current information on COVID-19 for healthcare personnel is available at CDC's webpage, *Information for Healthcare Professionals* (see links provided in "Where can I go for updates and more information" section).

What are the known and potential benefits and risks of using decontaminated N95 respirators?

Potential benefits include:

- May help prevent exposure to airborne pathogens, and therefore risk of infection or illness
- Extends the usability of compatible N95 respirators by allowing for decontamination and single-user reuse

- Availability of N95 respirators are critical to healthcare workers in the diagnosis and treatment of patients with COVID-19

Potential risks include:

- Failure of filtration efficiency
- Reduced breathability
- Strap failure and ineffective face-fit
- Reused respirators may not have been effectively decontaminated of SARS-CoV-2 or other pathogens

Overview of the ASP STERRAD Sterilization Systems

The ASP STERRAD Sterilization Systems, specifically the STERRAD 100S, NX, and 100NX Sterilization Systems, must be operated in the STERRAD 100S, NX Standard, and 100NX Express Cycles, respectively, to decontaminate compatible N95 respirators that are contaminated or potentially contaminated with SARS-CoV-2 and other pathogenic microorganisms so that the respirators can be reused by HCP. N95 or N95-equivalent respirators containing cellulose-based materials are not compatible with the ASP STERRAD Sterilization Systems.

The STERRAD Sterilization System decontaminates utilizing hydrogen peroxide vapor. The vaporized hydrogen peroxide is introduced to allow perfusion of the hydrogen peroxide throughout the chamber, facilitating hydrogen peroxide contact with the surfaces to be sterilized.

The ASP STERRAD Sterilization System enables single-user reuse of compatible N95 respirators that would otherwise be disposed of after a single use. However, respirators that are visibly soiled must be discarded and not reused or decontaminated.

What is an EUA?

The United States FDA has made the emergency use of the ASP STERRAD Sterilization System to decontaminate compatible N95 respirators available under an emergency access mechanism called an Emergency Use Authorization (EUA). The EUA is supported by the Secretary of Health and Human Service's (HHS's) declaration that circumstances exist to justify the

Report Adverse events, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 (<https://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home>) or by calling **1-800-FDA-1088**

FACT SHEET FOR HEALTHCARE PERSONNEL

ASP STERRAD Sterilization System for Decontaminating Compatible N95 Respirators April 11, 2020

Coronavirus
Disease 2019
(COVID-19)

emergency use of medical devices, including alternative products used as medical devices, due to insufficient supply during the COVID-19 pandemic. The ASP STERRAD Sterilization System for decontamination of compatible N95 respirators has been made available under an EUA, and has not undergone the same type of review as an FDA-approved or cleared device. FDA may issue an EUA when certain criteria are met, which includes that there are no adequate, approved, available alternatives, and based on the totality of scientific evidence available, it is reasonable to believe the ASP STERRAD Sterilization System may be effective at preventing exposure to pathogenic airborne particulates when there are insufficient supplies of respirators during the COVID-19 pandemic by decontaminating, for a maximum of 2 decontamination cycles per respirator, compatible N95 respirators that are contaminated or potentially contaminated with SARS-CoV-2 or other pathogenic microorganisms.

The EUA for the ASP STERRAD Sterilization System for decontamination of compatible N95 respirators is in effect for the duration of the COVID-19 declaration justifying emergency use of medical devices, unless terminated or revoked (after which the products may no longer be used).

Where can I go for updates and more information?

CDC webpages:

General: <https://www.cdc.gov/COVID19>

Healthcare Professionals:

<https://www.cdc.gov/coronavirus/2019-nCoV/guidance-hcp.html>

Infection Prevention and Control Recommendations in Healthcare Settings:

<https://www.cdc.gov/coronavirus/2019-ncov/infection-control/control-recommendations.html>

Infection Control: <https://www.cdc.gov/coronavirus/2019-ncov/infection-control/index.html>

FAQ on Personal Protective Equipment:

<https://www.cdc.gov/coronavirus/2019-ncov/hcp/respirator-use-faq.html>

FDA webpages:

General: www.fda.gov/novelcoronavirus

EUAs: <https://www.fda.gov/medical-devices/emergency-situations-medical-devices/emergency-use-authorizations>

Report Adverse events, including problems with test performance or results, to MedWatch by submitting the online FDA Form 3500 (<https://www.accessdata.fda.gov/scripts/medwatch/index.cfm?action=reporting.home>) or by calling **1-800-FDA-1088**